

Calibration Certificate No.28638

Instrument: **Microphone**
 Model: **KW4405**
 Manufacturer: **Audio Control**
 Serial number: **399**
 Part of: **CM-145 Microphone Unit**

Date Calibrated: **6/7/2013** Cal Due:
 Status:

Received	Sent
X	X

 In tolerance:

X	X
----------	----------

 Out of tolerance:

--	--

 See comments:

--	--

 Contains non-accredited tests: Yes No

Customer: **Studio Six Digital, LLC**
 Tel/Fax: **-**

Address: **1530 Sumac Ave.,
 Boulder CO 80304**



Tested in accordance with the following procedures and standards:

Calibration of Measurement Microphones, Scantek, Inc., Rev. 11/30/2010

Instrumentation used for calibration: N-1504 Norsonic Test System:

Instrument - Manufacturer	Description	S/N	Cal. Date	Traceability evidence	Cal. Due
				Cal. Lab / Accreditation	
483B-Norsonic	SME Cal Unit	31061	Jul 18, 2012	Scantek, Inc./ NVLAP	Jul 18, 2013
DS-360-SRS	Function Generator	88077	Aug 30, 2012	ACR Env./ A2LA	Aug 30, 2014
34401A-Agilent Technologies	Digital Voltmeter	MY47011118	Aug 28, 2012	ACR Env./ A2LA	Aug 28, 2013
HM30-Thommen	Meteo Station	1040170/39633	Dec 6, 2012	ACR Env./ A2LA	Dec 6, 2013
PC Program 1017 Norsonic	Calibration software	v.5.2	Validated Mar 2011	Scantek, Inc.	-
1253-Norsonic	Calibrator	31959	Nov 15, 2011	NPL-UK / UKAS	Nov 15, 2013
1203-Norsonic	Preamplifier	28969	Jun 5, 2013	Scantek, Inc./ NVLAP	Jun 5, 2014
4180-Brüel&Kjær	Microphone	2246115	Nov 21, 2011	NPL-UK / UKAS	Nov 21, 2013

Instrumentation and test results are traceable to SI - BIPM through standards maintained by NPL (UK) and NIST (USA)

Calibrated by:	Mariana Buzduga	Authorized signatory:	Valentin Buzduga
Signature		Signature	
Date	6/10/2013	Date	6/10/2013

Calibration Certificates or Test Reports shall not be reproduced, except in full, without written approval of the laboratory.
 This Calibration Certificate or Test Reports shall not be used to claim product certification, approval or endorsement by NVLAP, NIST, or any agency of the federal government.

Document stored as: Z:\Calibration Lab\Mic 2013\AuContCM145_399_M1.doc

Page 1 of 2